



LIMS Entry Summary

ARTG Number: 114907
ARTG Name: Poly Implant Protheses - High Cohesive Gel Pre-Filled Breast Implant (Textured) IMGHC-TX-S
Sponsor: Medical Vision Australia Pty Ltd
Batch Number: 25109

The sample was submitted by Sponsor

Results Approved By: Jorge Garcia **Sample Status:** VA

Results of the examination and analysis were as follows:

Approval Comments:

The sample is NOT cytotoxic

TEST	RESULT	COMMENTS	ENTERED BY	VALIDATED BY
CYTOTOXICITY med dev				
- S1	>100		wirasr	garcij
- S2	100		wirasr	garcij
- S3	>100		wirasr	garcij
- Positive control	13		wirasr	garcij
- Negative control	>100		wirasr	garcij
- Blank control	>100		wirasr	garcij
- Cytotoxic titre	100		wirasr	garcij