



Australian Government

Department of Health and Ageing
Therapeutic Goods Administration

LIMS Entry Summary

ARTG Number: 114908
ARTG Name: Poly Implant Protheses- High Cohesive Gel Pre-Filled Breast Implant (Smooth Texture) IMGHC-LS-S
Sponsor: Medical Vision Australia Pty Ltd
Batch Number: 56206

The sample was submitted by Sponsor

Results Approved By: Jorge Garcia **Sample Status:** VA

Results of the examination and analysis were as follows:

Approval Comments:
The Sample is NOT cytotoxic

TEST	RESULT	COMMENTS	ENTERED BY	VALIDATED BY
CYTOTOXICITY med dev				
- S1	>100		carsca	garcej
- S2	>100		carsca	garcej
- S3	>100		carsca	garcej
- Positive control	20		carsca	garcej
- Negative control	>100		carsca	garcej
- Blank control	>100		carsca	garcej
- Cytotoxic titre	>100		carsca	garcej