



**Australian Government**

**Department of Health and Ageing**  
Therapeutic Goods Administration

**LIMS Entry Summary**

**ARTG Number:** 114900  
**ARTG Name:** Poly Implant Protheses - High Cohesive Gel Pre-Filled Breast Implant (Smooth Texture) IMGHC-LS-H  
**Sponsor:** Medical Vision Australia Pty Ltd  
**Batch Number:** 36709

The sample was submitted by Sponsor

**Results Approved By:** Jorge Garcia      **Sample Status:** VA  
 Results of the examination and analysis were as follows:

**Approval Comments:**  
The sample is NOT cytotoxic

TEST	RESULT	COMMENTS	ENTERED BY	VALIDATED BY
CYTOTOXICITY med dev				
- S1	>100		carsca	garcij
- S2	>100		carsca	garcij
- S3	>100		carsca	garcij
- Positive control	13		carsca	garcij
- Negative control	>100		carsca	garcij
- Blank control	>100		carsca	garcij
- Cytotoxic titre	>100		carsca	garcij