

## PUBLIC SUMMARY DOCUMENT

**Product:** Hollister New Image Urostomy

**Applicant:** Hollister

**Date of SPAP Meeting:** 22 October 2019

### 1. Proposed Variation on the Stoma Appliance Scheme

The applicant, Hollister, sought a change in product description for three existing variants to the current listing of the Hollister New Image Urostomy (SAS Code 80018L) in subgroup 7(a) of the Stoma Appliance Scheme (SAS) Schedule. The product, including six variants, is currently listed at a unit price of \$3.504, with a maximum monthly quantity of 60 units.

### 2. Variants to be added

The applicant advised the following pouch film changes to three existing variants for SAS Code 80018L.

Product Code	Description
18902	Non-return valve, Multi Chamber, transparent with fabric backing, Large, 590ml, standard headspace, direct connection to drainage bag, tap closure, belt loops on each side of pouch, mechanical coupling, 44mm ring <u>Pouch film changes</u> Chlorine free, white pouch film replaced by beige film, translucent pouch film replaced by a more transparent film
18903	Non-return valve with Multi Chamber, transparent with fabric backing, large, 590ml, standard headspace, direct connection to drainage bag, tap closure, belt loops on each side of pouch, mechanical coupling, 57mm ring <u>Pouch film changes</u> Chlorine free, white pouch film replaced by beige film, translucent pouch film replaced by a more transparent film
18904	Non-return valve with Multi Chamber, transparent with fabric backing, large, 590ml, standard headspace, direct connection to drainage bag, tap closure, belt loops on each side of pouch, mechanical coupling, 70mm ring <u>Pouch film changes</u> Chlorine free, white pouch film replaced by beige film, translucent pouch film replaced by a more transparent film

### 3. Background

This product was first listed on the SAS Schedule in May 2015.

### 4. Clinical Place for the Product

The proposed product provides an alternative for users requiring a Two-Piece Urostomy with Mechanical Coupling.

### *Financial Analysis*

Not applicable

## **5. SPAP Recommendation**

The SPAP recommended the change in pouch film to three existing variants of the current product range for Hollister New Image Urostomy (SAS Code 80018L) listed in subgroup 7(a) of the SAS Schedule, at the unit price of \$3.504, with a maximum monthly quantity of 60 units.

## **6. Context for Decision**

The SPAP helps decide whether stoma products should be subsidised and, if so, the conditions of their subsidisation in Australia. It considers submissions in this context. An SPAP decision not to recommend listing or changes to a listing does not represent a final SPAP view about the merits of a particular stoma product. A company can resubmit to the SPAP following a decision not to recommend listing or changes to a listing. The SPAP is an advisory committee and as such its recommendations are non-binding on Government. All SPAP recommendations are subject to Cabinet/Ministerial approval.

## **7. Applicant's Comment**

Hollister accepts the SPAP's recommendation.