

PUBLIC SUMMARY DOCUMENT

Product: Omnigon Flexima One Piece Closed

Applicant: Omnigon

Date of Meeting: Out-of-Session November 2020

1. Proposed Variation on the Stoma Appliance Scheme

The applicant, Omnigon, sought improvements/changes for two existing products (44070 and 44170) of the current listing of the Omnigon Flexima (SAS Code 9926K) in subgroup 1(b) of the Stoma Appliance Scheme (SAS) Schedule. The product, including 10 variants, is currently listed at a unit price of \$2.735, with a maximum monthly quantity of 90 units.

2. Variants to be added

The applicant advised the following improvements/changes for two existing products of SAS code 9926K.

Product Code	Improvements/Changes
44070	New print on the white backing paper now has CTF 15-75mm, previously CTF 15-70mm, an increase of 5mm. New ergonomic round shape for increased comfort, pouch volume is unchanged. New integrated high-protected (HP) filter with active charcoal, protected by a breathable membrane. This is the same filter as the Flexima Active closed flat pouches (SAS Code 80021P). New inspection window.
44170	New print on the white backing paper now has CTF 15-75mm, previously CTF 15-70mm, an increase of 5mm. New ergonomic round shape for increased comfort, pouch volume is unchanged. New integrated high-protected (HP) filter with active charcoal, protected by a breathable membrane. This is the same filter as the Flexima Active closed flat pouches (SAS Code 80021P).

3. Background

This product was first listed on the SAS Schedule in April 2011.

4. Clinical Place for the Product

The proposed product provides an alternative for users requiring a one-piece closed pouch.

Financial Analysis

Not applicable.

5. SPAP Recommendation

The Panel recommended the improvements/changes for the two existing products (44070 and 44170) of the current range of Omnigon Flexima (SAS Code 9926K), listed in subgroup 1(b), of the SAS Schedule, at the unit price of \$2.735, with a maximum monthly quantity of 90 units.

6. Context for Decision

The Panel helps decide whether stoma products should be subsidised and, if so, the conditions of their subsidisation in Australia. It considers submissions in this context. A Panel decision not to recommend listing or changes to a listing does not represent a final Panel view about the merits of a particular stoma product. A company can resubmit to the Panel following a decision not to recommend listing or changes to a listing. The Panel is an advisory committee and as such its recommendations are non-binding on Government. All Panel recommendations are subject to Cabinet/Ministerial approval.

7. Applicant's Comment

Omnigon agrees with the Panel's recommendation.